

Corticosterone ELISA Kit

Item No. 501320



Customer Service 800.364.9897 * Technical Support 888.526.5351

www.caymanchem.com

TABLE OF CONTENTS

GENERAL INFORMATION	3	Materials Supplied
	4	Safety Data
	4	Precautions
	4	If You Have Problems
	5	Storage and Stability
	5	Materials Needed but Not Supplied
INTRODUCTION	6	About This Assay
	7	Description of AChE Competitive ELISAs
	9	Biochemistry of Acetylcholinesterase
	11	Definition of Key Terms
PRE-ASSAY PREPARATION	12	Buffer Preparation
	13	Sample Preparation
	14	Sample Purification
ASSAY PROTOCOL	20	Preparation of Assay-Specific Reagents
	22	Plate Set Up
	23	Performing the Assay
ANALYSIS	26	Calculations
	28	Performance Characteristics
RESOURCES	32	Troubleshooting
	33	References
	35	Plate Template
	36	Notes
	36	Warranty and Limitation of Remedy

GENERAL INFORMATION

Materials Supplied

Item Number	Item	96 wells Quantity/Size	480 wells Quantity/Size
401322	Corticosterone ELISA Antiserum	1 vial/100 dtn	1 vial/500 dtn
401320	Corticosterone AChE Tracer	1 vial/100 dtn	1 vial/500 dtn
400657	Corticosterone ELISA Standard	1 vial	1 vial
400060	ELISA Buffer Concentrate (10X)	2 vials/10 ml	4 vials/10 ml
400062	Wash Buffer Concentrate (400X)	1 vial/5 ml	1 vial/12.5 ml
400035	Polysorbate 20	1 vial/3 ml	1 vial/3 ml
400004/400006	Mouse Anti-Rabbit IgG Coated Plate	1 plate	5 plates
400012	96-Well Cover Sheet	1 cover	5 covers
400050	Ellman's Reagent	3 vials/100 dtn	6 vials/250 dtn

If any of the items listed above are damaged or missing, please contact our Customer Service department at (800) 364-9897 or (734) 971-3335. We cannot accept any returns without prior authorization.



WARNING: THIS PRODUCT IS FOR RESEARCH ONLY - NOT FOR HUMAN OR VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

Safety Data

This material should be considered hazardous until further information becomes available. Do not ingest, inhale, get in eyes, on skin, or on clothing. Wash thoroughly after handling. Before use, the user must review the complete Safety Data Sheet, which has been sent *via* email to your institution.

Precautions

Please read these instructions carefully before beginning this assay.

The reagents in this kit have been tested and formulated to work exclusively with Cayman Chemical's AChE ELISA Kits. This kit may not perform as described if any reagent or procedure is replaced or modified.

When compared to quantification by LC/MS or GC/MS, it is not uncommon for immunoassays to report higher analyte concentrations. While LC/MS or GC/MS analyses typically measure only a single compound, antibodies used in immunoassays sometimes recognize not only the target molecule, but also structurally related molecules, including biologically relevant metabolites. In many cases, measurement of both the parent molecule and metabolites is more representative of the overall biological response than is the measurement of a short-lived parent molecule. It is the responsibility of the researcher to understand the limits of both assay systems and to interpret their data accordingly.

If You Have Problems

Technical Service Contact Information

Phone: 888-526-5351 (USA and Canada only) or 734-975-3888

Fax: 734-971-3641

Email: techserv@caymanchem.com

Hours: M-F 8:00 AM to 5:30 PM EST

In order for our staff to assist you quickly and efficiently, please be ready to supply the lot number of the kit (found on the outside of the box).

Storage and Stability

This kit will perform as specified if stored as directed at -20°C and used before the expiration date indicated on the outside of the box.

Materials Needed But Not Supplied

1. A plate reader capable of measuring absorbance between 405-420 nm.
2. Adjustable pipettes and a repeating pipettor.
3. A source of 'UltraPure' water. Water used to prepare all ELISA reagents and buffers must be deionized and free of trace organic contaminants ('UltraPure'). Use activated carbon filter cartridges or other organic scavengers. Glass distilled water (even if double distilled), HPLC-grade water, and sterile water (for injections) are not adequate for ELISA. *NOTE: UltraPure water is available for purchase from Cayman (Item No. 400000).*
4. Materials used for **Sample Purification** (see page 13).

About This Assay

Corticosterone is a steroid hormone produced by the adrenal cortex in response to stress. The production of glucocorticoids is increased in stress, and corticosterone is frequently measured as an indicator of stress.¹ Corticosterone can be measured in a variety of sample matrices including plasma, serum, feces, urine, eggs, and feathers.¹⁻⁸

Cayman's Corticosterone ELISA Kit is a competitive assay that has been developed for the measurement of corticosterone in serum, feces and other sample matrices. The assay has a range of 8.2-5,000 pg/ml and a sensitivity (80% B/B₀) of approximately 30 pg/ml.

Description of AChE Competitive ELISAs^{11,12}

This assay is based on the competition between corticosterone and a corticosterone-acetylcholinesterase (AChE) conjugate (Corticosterone Tracer) for a limited amount of Corticosterone Antiserum. Because the concentration of the Corticosterone Tracer is held constant while the concentration of corticosterone varies, the amount of Corticosterone Tracer that is able to bind to the Corticosterone Antiserum will be inversely proportional to the concentration of corticosterone in the well. This antiserum-corticosterone (either free or tracer) complex binds to the mouse anti-rabbit IgG that has been previously attached to the well. The plate is washed to remove any unbound reagents and then Ellman's Reagent (which contains the substrate to AChE) is added to the well. The product of this enzymatic reaction has a distinct yellow color and absorbs strongly at 412 nm. The intensity of this color, determined spectrophotometrically, is proportional to the amount of Corticosterone Tracer bound to the well, which is inversely proportional to the amount of free corticosterone present in the well during the incubation; or

$$\text{Absorbance} \propto [\text{Bound Corticosterone Tracer}] \propto 1/[\text{Corticosterone}]$$

A schematic of this process is shown in Figure 1, on page 8.




Figure 1. Schematic of the AChE ELISA

Biochemistry of Acetylcholinesterase

The electric organ of the electric eel, *E. electricus*, contains an avid AChE capable of massive catalytic turnover during the generation of its electrochemical discharges. The electric eel AChE has a clover leaf-shaped tertiary structure consisting of a triad of tetramers attached to a collagen-like structural fibril. This stable enzyme is capable of high turnover ($64,000 \text{ s}^{-1}$) for the hydrolysis of acetylthiocholine.

A molecule of the analyte covalently attached to a molecule of AChE serves as the tracer in AChE enzyme immunoassays. Quantification of the tracer is achieved by measuring its AChE activity with Ellman's Reagent. This reagent consists of acetylthiocholine and 5,5'-dithio-bis-(2-nitrobenzoic acid). Hydrolysis of acetylthiocholine by AChE produces thiocholine (see Figure 2, on page 10). The non-enzymatic reaction of thiocholine with 5,5'-dithio-bis-(2-nitrobenzoic acid) produces 5-thio-2-nitrobenzoic acid, which has a strong absorbance at 412 nm ($\epsilon = 13,600$).

AChE has several advantages over other enzymes commonly used for enzyme-linked immunosorbent assays. Unlike horseradish peroxidase, AChE does not self-inactivate during turnover. This property of AChE also allows redevelopment of the assay if it is accidentally splashed or spilled. In addition, the enzyme is highly stable under the assay conditions, has a wide pH range (pH 5-10), and is not inhibited by common buffer salts or preservatives. Since AChE is stable during the development step, it is unnecessary to use a 'stop' reagent, and the plate may be read whenever it is convenient.




Figure 2. Reaction catalyzed by acetylcholinesterase

Definition of Key Terms

Blank: background absorbance caused by Ellman's Reagent. The blank absorbance should be subtracted from the absorbance readings of all the other wells, including NSB wells.

Total Activity: total enzymatic activity of the AChE-linked tracer. This is analogous to the specific activity of a radioactive tracer.

NSB (Non-Specific Binding): non-immunological binding of the tracer to the well. Even in the absence of specific antibody a very small amount of tracer still binds to the well; the NSB is a measure of this low binding. Do not forget to subtract the Blank absorbance values.

B_0 (Maximum Binding): maximum amount of the tracer that the antibody can bind in the absence of free analyte.

$\%B/B_0$ (%Bound/Maximum Bound): ratio of the absorbance of a particular sample or standard well to that of the maximum binding (B_0) well.

Standard Curve: a plot of the $\%B/B_0$ values *versus* concentration of a series of wells containing various known amounts of analyte.

Dtn: determination, where one dtn is the amount of reagent used per well.

PRE-ASSAY PREPARATION

NOTE: Water used to prepare all ELISA reagents and buffers must be deionized and free of trace organic contaminants ('UltraPure'). Use activated carbon filter cartridges or other organic scavengers. Glass distilled water (even if double distilled), HPLC-grade water, and sterile water (for injections) are not adequate for ELISA. UltraPure water may be purchased from Cayman (Item No. 400000).

Buffer Preparation

Store all diluted buffers at 4°C; they will be stable for about two months.

1. ELISA Buffer Preparation

Dilute the contents of one vial of ELISA Buffer Concentrate (10X) (Item No. 400060) with 90 ml of UltraPure water. Be certain to rinse the vial to remove any salts that may have precipitated. *NOTE: It is normal for the concentrated buffer to contain crystalline salts after thawing. These will completely dissolve upon dilution with water.*

2. Wash Buffer Preparation

5 ml vial Wash Buffer Concentrate (400X) (96-well kit; Item No. 400062): Dilute to a total volume of 2 liters with UltraPure water and add 1 ml of Polysorbate 20 (Item No. 400035).

OR

12.5 ml vial Wash Buffer Concentrate (400X) (480-well kit; Item No. 400062): Dilute to a total volume of 5 liters with UltraPure water and add 2.5 ml of Polysorbate 20 (Item No. 400035).

Smaller volumes of Wash Buffer can be prepared by diluting the Wash Buffer Concentrate 1:400 and adding Polysorbate 20 (0.5 ml/liter of Wash Buffer).

NOTE: Polysorbate 20 is a viscous liquid and cannot be measured by a regular pipette. A positive displacement pipette or a syringe should be used to deliver small quantities accurately.

Sample Preparation

This assay has been validated for rodent serum and feces. Other sample types should be checked for interference before embarking on a large number of sample measurements. To test for interference, dilute one or two test samples to obtain at least two different dilutions of each sample between approximately 30 and 1,500 pg/ml (i.e., between ~20-80% B/B₀). If the two different dilutions of the sample show good correlations (differ by 20% or less) in the final calculated corticosterone concentration, purification is not required. If you do not see good correlation of the different dilutions, purification is advised. Proper sample storage and preparation are essential for consistent and accurate results. Please read this section thoroughly before beginning the assay.

General Precautions

- All samples must be free of organic solvents prior to assay.
- Samples should be assayed immediately after collection; samples that cannot be assayed immediately should be stored at -80°C.
- Samples of rabbit origin may contain antibodies which interfere with the assay by binding to the mouse anti-rabbit IgG plate. We recommend that all sheep and rabbit samples be purified prior to use in this assay.
- AEBSF (Pefabloc SC[®]) and PMSF inhibit AChE. Samples containing these protease inhibitors should not be used in this assay.

Serum

Collect blood in vacutainers that do not contain anticoagulant. Allow samples to clot undisturbed for 30-60 minutes. Remove the clot by centrifugation at 1-2,000 x g for 15 minutes. Samples should be assayed immediately after collection; samples that cannot be assayed immediately should be stored at -20°C. Process following the **Sample Purification** methods beginning on page 14.

Feces

NOTE: This kit has been validated for use with mouse fecal samples. The antiserum in this kit does not recognize the fecal metabolites of corticosterone found in deer and elk. If you wish to use this kit for fecal material from other species, it is strongly recommended that you perform a biological validation before proceeding with the analysis of a large number of samples.

Fecal samples should be frozen immediately after collection. Process following the Fecal Sample Extraction methods beginning on page 18.

Sample Purification

Purification Protocol

Materials Needed

1. Tritium-labeled corticosterone (optional)
2. Methylene chloride (for plasma samples)
3. 80% Methanol (for fecal samples)

Serum Sample Extraction

The following protocol is a suggestion only. You may choose a different protocol based on your own requirements, sample type, and expertise. If desired, recovery may be tracked by spiking samples with tritium-labeled corticosterone ($[^3\text{H}]$ -corticosterone) and follow the **Spiked-Sample Recovery Calculations** in the **Analysis** section on page 27. Otherwise, omit step 2.

NOTE: We do not recommend the use of plastic vials or caps for this procedure. The methylene chloride may extract interfering compounds from the plastic.

1. Aliquot a known amount of each sample into a clean test tube.
2. Add 10,000 cpm of $[^3\text{H}]$ -corticosterone. Use a high specific activity tracer to minimize the amount of radioactive cortico-sterone as the ELISA will be able to detect the added corticosterone. Follow the procedure below for both spiked and unspiked samples.
3. Add methylene chloride (approximately four times the sample volume) to each tube. Vortex to mix thoroughly. Allow the layers to separate. Transfer the methylene chloride (lower) layer into a clean test tube using a transfer pipette. Repeat this step three times.

NOTE: If it is necessary to stop during this purification, samples may be stored in the methylene chloride solution at -20°C or -80°C .

4. Evaporate the methylene chloride under a gentle stream of nitrogen.
5. Dissolve the extract in ELISA Buffer. If the samples were spiked with $[^3\text{H}]$ -corticosterone to track recovery, remove 10% of the resuspended volume for scintillation counting. The remainder is available for use in the ELISA.

Level	Average (ng/ml)	%CV Intra-assay variation	Average (ng/ml)	%CV Inter-assay variation
High	654	9.3	714	7.0
Medium	372	7.4	406	7.0
Low	123	4.9	147	8.8

Table 1. Mouse serum sample validation

Mouse serum samples containing a high, medium, or low level of corticosterone were measured 60 times each using a single set of reagents. The calculated %CV is reported as intra-assay variance. A separate series of mouse serum samples containing a high, medium, or low level of corticosterone were measured four times each using eight independent sets of reagents. The calculated %CV is reported as inter-assay variance.




Figure 3. Recovery of corticosterone from mouse serum

Mouse serum samples were spiked with corticosterone, purified as described in the Sample Preparation and Sample Purification sections and analyzed using the Corticosterone ELISA Kit. The y-intercept corresponds to the amount of corticosterone in unspiked mouse serum. Error bars represent standard deviations obtained from multiple dilutions of each sample.

Fecal Sample Extraction

The following protocol is a suggestion only. You may choose a different protocol based on your own requirements and expertise. *NOTE: It is important to transfer supernatants to clean tubes immediately after completing the extraction procedure, as some fecal metabolites will continue to elute into the methanol as long as there is contact between the two materials.*

1. Lyophilize samples to remove water.
2. Sift lyophilized samples through a stainless steel mesh to remove large particles. Mix each sample thoroughly to ensure homogeneity.
3. Place 50 mg of each sample into clean test tubes.
4. Add 1 ml of 80% methanol to each sample.
5. Vortex at high speed for thirty minutes.
6. Centrifuge at 2,500 x g for 20 minutes.
7. Transfer each supernatant to a clean test tube.
8. Evaporate methanol under a gentle stream of nitrogen. Alternatively, samples can be measured without drying provided that they will be diluted at least 1:50 in ELISA Buffer prior to assay.
9. Suspend dried samples in 500 µl of ELISA Buffer. The samples are now ready for use in the ELISA.

Level	Average (pg/mg)	%CV Intra-assay variation	Average (pg/mg)	%CV Inter-assay variation
High	723	3.8	739	6.7
Medium	271	4.8	286	8.6
Low	178	5.4	201	9.8

Table 2. Mouse fecal sample validation

Mouse fecal samples containing a high, medium, or low level of corticosterone were measured 60 times each using a single set of reagents. The calculated %CV is reported as intra-assay variance. A separate series of mouse fecal samples containing a high, medium, or low level of corticosterone were measured four times each using eight independent sets of reagents. The calculated %CV is reported as inter-assay variance.

Preparation of Assay-Specific Reagents

Corticosterone ELISA Standard

Equilibrate a pipette tip in ethanol by repeatedly filling and expelling the tip with ethanol several times. Using the equilibrated pipette tip, transfer 100 μ l of the Corticosterone ELISA Standard (Item No. 400657) into a clean test tube, then dilute with 900 μ l UltraPure water. The concentration of this solution (the bulk standard) will be 50 ng/ml. Store this solution at 4°C; it will be stable for at least six weeks.

NOTE: If assaying culture medium samples, culture medium should be used in place of ELISA Buffer for dilution of the standard curve and samples.

To prepare the standard for use in ELISA: obtain eight clean test tubes and number them #1 through #8. Aliquot 900 μ l ELISA Buffer to tube #1 and 750 μ l ELISA Buffer to tubes #2-8. Transfer 100 μ l of the bulk standard (50 ng/ml) to tube #1 and mix thoroughly. Serially dilute the standard by removing 500 μ l from tube #1 and placing it into tube #2; mix thoroughly. Next, remove 500 μ l from tube #2 and place it into tube #3; mix thoroughly. Repeat this process for tubes #4-8. These diluted standards should not be stored for more than 24 hours.




Figure 4. Preparation of the corticosterone standards

Corticosterone AChE Tracer

Reconstitute the Corticosterone AChE Tracer as follows:

100 dtn Corticosterone AChE Tracer (96-well kit; Item No. 401320):
Reconstitute with 6 ml ELISA Buffer.

OR

500 dtn Corticosterone AChE Tracer (480-well kit; Item No. 401320):
Reconstitute with 30 ml ELISA Buffer.

Store the reconstituted Corticosterone AChE Tracer at 4°C (*do not freeze!*) and use within four weeks. A 20% surplus of tracer has been included to account for any incidental losses.

Corticosterone ELISA Antiserum

Reconstitute the Corticosterone ELISA Antiserum as follows:

100 dtn Corticosterone ELISA Antiserum (96-well kit; Item No. 401322):
Reconstitute with 6 ml ELISA Buffer.

OR

500 dtn Corticosterone ELISA Antiserum (480-well kit; Item No. 401322):
Reconstitute with 30 ml ELISA Buffer.

Store the reconstituted Corticosterone ELISA Antiserum at 4°C. It will be stable for at least four weeks. A 20% surplus of antiserum has been included to account for any incidental losses.

Plate Set Up

The 96-well plate(s) included with this kit is supplied ready to use. It is not necessary to rinse the plate(s) prior to adding the reagents. *NOTE: If you do not need to use all the strips at once, place the unused strips back in the plate packet and store at 4°C. Be sure the packet is sealed with the desiccant inside.*

Each plate or set of strips must contain a minimum of two blanks (Blks), two non-specific binding wells (NSBs), two maximum binding wells (B_0 s), and an eight point standard curve run in duplicate. *NOTE: Each assay must contain this minimum configuration in order to ensure accurate and reproducible results.* Each sample should be assayed at two dilutions and each dilution should be assayed in duplicate. For statistical purposes, we recommend assaying samples in triplicate.

A suggested plate format is shown in Figure 5, below. The user may vary the location and type of wells present as necessary for each particular experiment. The plate format provided below has been designed to allow for easy data analysis using a convenient spreadsheet offered by Cayman (on page 26, for more details). We suggest you record the contents of each well on the template sheet provided (on page 35).

	1	2	3	4	5	6	7	8	9	10	11	12
A	Blk	S1	S1	1	1	1	9	9	9	17	17	17
B	Blk	S2	S2	2	2	2	10	10	10	18	18	18
C	NSB	S3	S3	3	3	3	11	11	11	19	19	19
D	NSB	S4	S4	4	4	4	12	12	12	20	20	20
E	B_0	S5	S5	5	5	5	13	13	13	21	21	21
F	B_0	S6	S6	6	6	6	14	14	14	22	22	22
G	B_0	S7	S7	7	7	7	15	15	15	23	23	23
H	TA	S8	S8	8	8	8	16	16	16	24	24	24

Blk - Blank
TA - Total Activity
NSB - Non-Specific Binding
 B_0 - Maximum Binding
S1-S8 - Standards 1-8
1-24 - Samples

Figure 5. Sample plate format

Performing the Assay

Pipetting Hints

- Use different tips to pipette each reagent.
- Before pipetting each reagent, equilibrate the pipette tip in that reagent (i.e., slowly fill the tip and gently expel the contents, repeat several times).
- Do not expose the pipette tip to the reagent(s) already in the well(s).

Addition of the Reagents

1. ELISA Buffer

Add 100 μ l ELISA Buffer to NSB wells. Add 50 μ l ELISA Buffer to B_0 wells. If culture medium was used to dilute the standard curve, substitute 50 μ l of culture medium for ELISA Buffer in the NSB and B_0 wells (i.e., add 50 μ l culture medium to NSB and B_0 wells and 50 μ l ELISA Buffer to NSB wells).

2. Corticosterone ELISA Standard

Add 50 μ l from tube #8 to both of the lowest standard wells (S8). Add 50 μ l from tube #7 to each of the next two standard wells (S7). Continue with this procedure until all the standards are aliquoted. The same pipette tip should be used to aliquot all the standards. Before pipetting each standard, be sure to equilibrate the pipette tip in that standard.

3. Samples

Add 50 μ l of sample per well. Each sample should be assayed at a minimum of two dilutions. Each dilution should be assayed in duplicate (triplicate recommended).

4. Corticosterone AChE Tracer

Add 50 μ l to each well *except* the TA and the Blk wells.

5. Corticosterone ELISA Antiserum

Add 50 μ l to each well *except* the TA, the NSB, and the Blk wells.

Well	ELISA Buffer	Standard/ Sample	Tracer	Antiserum
Blk	-	-	-	-
TA	-	-	5 µl (at devel. step)	-
NSB	100 µl	-	50 µl	-
B ₀	50 µl	-	50 µl	50 µl
Std/Sample	-	50 µl	50 µl	50 µl

Table 3. Pipetting summary

Incubation of the Plate

Cover each plate with plastic film (Item No. 400012) and incubate overnight at 4°C.

Development of the Plate

1. Reconstitute Ellman's Reagent immediately before use (20 ml of reagent is sufficient to develop 100 wells):

100 dtn vial Ellman's Reagent (96-well kit; Item No. 400050): Reconstitute with 20 ml of UltraPure water.

OR

250 dtn vial Ellman's Reagent (480-well kit; Item No. 400050): Reconstitute with 50 ml of UltraPure water.

NOTE: Reconstituted Ellman's Reagent is unstable and should be used the same day it is prepared; protect the Ellman's Reagent from light when not in use. Extra vials of the reagent have been provided should a plate need to be re-developed or multiple assays be run on different days.

2. Empty the wells and rinse five times with Wash Buffer.
3. Add 200 µl of Ellman's Reagent to each well.
4. Add 5 µl of tracer to the TA well.
5. Cover the plate with plastic film. Optimum development is obtained by using an orbital shaker equipped with a large, flat cover to allow the plate(s) to develop in the dark. This assay typically develops (*i.e.*, B₀ wells ≥0.3 A.U. (blank subtracted)) in 90-120 minutes.

Reading the Plate

1. Wipe the bottom of the plate with a clean tissue to remove fingerprints, dirt, etc.
2. Remove the plate cover being careful to keep Ellman's Reagent from splashing on the cover. *NOTE: Any loss of Ellman's Reagent will affect the absorbance readings. If Ellman's Reagent is present on the cover, use a pipette to transfer the Ellman's Reagent into the well. If too much Ellman's Reagent has splashed on the cover to easily redistribute back into the wells, wash the plate three times with wash buffer and repeat the development with fresh Ellman's Reagent.*
3. Read the plate at a wavelength between 405 and 420 nm (usually 412 nm). The absorbance may be checked periodically until the B₀ wells have reached a minimum of 0.3 A.U. (blank subtracted). The plate should be read when the absorbance of the B₀ wells are in the range of 0.3-1.5 A.U. (blank subtracted). If the absorbance of the wells exceeds 2.0, wash the plate, add fresh Ellman's Reagent and let it develop again.

ANALYSIS

Many plate readers come with data reduction software that plot data automatically. Alternatively a spreadsheet program can be used. The data should be plotted as either %B/B₀ versus log concentration using a four-parameter logistic fit or as logit B/B₀ versus log concentration using a linear fit. NOTE: Cayman has a computer spreadsheet available for data analysis. Please contact Technical Service or visit our website (www.caymanchem.com/analysis/elisa) to obtain a free copy of this convenient data analysis tool.

Calculations

Preparation of the Data

The following procedure is recommended for preparation of the data prior to graphical analysis.

NOTE: If the plate reader has not subtracted the absorbance readings of the blank wells from the absorbance readings of the rest of the plate, be sure to do that now.

1. Average the absorbance readings from the NSB wells.
2. Average the absorbance readings from the B₀ wells.
3. Subtract the NSB average from the B₀ average. This is the corrected B₀ or corrected maximum binding.
4. Calculate the B/B₀ (Sample or Standard Bound/Maximum Bound) for the remaining wells. To do this, subtract the average NSB absorbance from the S1 absorbance and divide by the corrected B₀ (from Step 3). Repeat for S2-S8 and all sample wells. (To obtain %B/B₀ for a logistic four-parameter fit, multiply these values by 100.)

NOTE: The TA values are not used in the standard curve calculations. Rather, they are used as a diagnostic tool; the corrected B₀ divided by the actual TA (10X measured absorbance) will give the % Bound. This value should closely approximate the % Bound that can be calculated from the **Sample Data** (see page 28). Erratic absorbance values and a low (or no) % Bound could indicate the presence of organic solvents in the buffer or other technical problems (see page 32 for **Troubleshooting**).

Plot the Standard Curve

Plot %B/B₀ for standards S1-S8 versus corticosterone concentration (usually in pg/ml) using linear (y) and log (x) axes and perform a 4-parameter logistic fit.

Alternative Plot - The data can also be linearized using a logit transformation. The equation for this conversion is shown below. NOTE: Do not use %B/B₀ in this calculation.

$$\text{logit (B/B}_0\text{)} = \ln [B/B_0/(1 - B/B_0)]$$

Plot the data as logit (B/B₀) versus log concentrations and perform a linear regression fit.

Determine the Sample Concentration

Calculate the B/B₀ (or %B/B₀) value for each sample. Determine the concentration of each sample using the equation obtained from the standard curve plot. NOTE: Remember to account for any concentration or dilution of the sample prior to the addition to the well. Samples with %B/B₀ values greater than 80% or less than 20% should be re-assayed as they generally fall out of the linear range of the standard curve. A 20% or greater disparity between the apparent concentration of two different dilutions of the same sample indicates interference which could be eliminated by purification.

Spiked-Sample Recovery Calculation

$$\text{Recovery Factor} = \frac{10X \text{ cpm of sample}}{[^3\text{H}]\text{-Corticosterone added to sample}}$$

Corticosterone (pg) in purified sample =

$$\left[\frac{\text{Value from ELISA (pg/ml)}}{\text{Recovery Factor}} \right] \times \text{reconstituted volume of sample} - \text{added } [^3\text{H}]\text{-Corticosterone (pg)}$$

$$\text{Corticosterone in sample (pg/ml)} = \frac{\text{Corticosterone (pg) in purified sample}}{\text{Volume of sample used for purification (ml)}}$$

Performance Characteristics


Sample Data

The standard curve presented here is an example of the data typically produced with this kit; however, your results will not be identical to these. You **must** run a new standard curve. Do not use the data below to determine the values of your samples. Your results could differ substantially.

	Raw Data		Average	Corrected
Total Activity	0.419	0.351	0.385	
NSB	0.001	0.001	0.001	
B_0	1.456	1.463		
	1.476	1.445	1.460	1.459

Dose (pg/ml)	Raw Data		Corrected		%B/ B_0	
5,000	0.099	0.107	0.098	0.106	6.7	7.2
2,000	0.184	0.187	0.183	0.186	12.5	12.7
800	0.329	0.333	0.328	0.332	22.4	22.7
320	0.528	0.525	0.527	0.524	36.1	35.9
128	0.777	0.769	0.776	0.768	53.2	52.6
51.2	1.014	1.003	1.013	1.002	69.4	68.7
20.5	1.192	1.183	1.191	1.182	81.7	81.0
8.2	1.301	1.307	1.300	1.306	89.1	89.5

Table 4. Typical results



Assay Range = 8.2-5,000 pg/ml
Sensitivity (defined as 80% B/ B_0) = ~30 pg/ml
Mid-point (defined as 50% B/ B_0) = ~130-180 pg/ml

The sensitivity and mid-point were derived from the standard curve shown above. The standard was diluted with ELISA Buffer.

Figure 6. Typical standard curve

Precision:

The intra-assay CVs have been determined at multiple points on the standard curve. These data are summarized in the graph on page 29 and in the table below.

Dose (pg/ml)	%CV* Intra-assay variation
5,000	6.0
2,000	2.1
800	2.7
320	4.3
128	4.9
51.2	5.7
20.5	11.6
8.2	12.3

Table 5. Intra-assay variation

*%CV represents the variation in concentration (not absorbance) as determined using a reference standard curve.

Specificity:

Compound	Cross Reactivity
Corticosterone	100%
11-Deoxycorticosterone	15.8%
Prednisolone	3.4%
11-Dehydrocorticosterone	2.9%
Cortisol	2.5%
Progesterone	1.4%
Aldosterone	0.47%
17 α -Hydroxyprogesterone	0.21%
11-Deoxycortisol	0.14%
Androstenedione	0.11%
Testosterone	0.07%
Pregnenolone	0.03%
DHEA sulfate	0.0005%
5 α -DHT	<0.01%
17 α -Hydroxypregnenolone	<0.01%
Androstenediol	<0.01%
DHEA	<0.01%
Estradiol	<0.01%
Estriol	<0.01%
Estrone sulfate	<0.01%
Ethinylestradiol	<0.01%
Dexamethasone	<0.01%

Table 6. Specificity of the Corticosterone ELISA Antiserum

RESOURCES

Troubleshooting

Problem	Possible Causes	Recommended Solutions
Erratic values; dispersion of duplicates	A. Trace organic contaminants in the water source B. Poor pipetting/technique	A. Replace activated carbon filter or change source of UltraPure water
High NSB (>10% of B ₀)	A. Poor washing B. Exposure of NSB wells to specific antibody	A. Re-wash plate and redevelop
Very low B ₀	A. Trace organic contaminants in the water source B. Plate requires additional development time C. Dilution error in preparing reagents	A. Replace activated carbon filter or change source of UltraPure water B. Return plate to shaker and re-read later
Low sensitivity (shift in dose response curve)	Standard is degraded	Replace standard
Analyses of two dilutions of a biological sample do not agree (i.e., more than 20% difference)	Interfering substances are present	Purify sample prior to analysis by ELISA ¹³
Only Total Activity (TA) wells develop	Trace organic contaminants in the water source	Replace activated carbon filter or change source of UltraPure water

References

- Goymann, W., Möstl, E., and Gwinner, E. Corticosterone metabolites can be measured noninvasively in excreta of European stonechats (*Saxicola torquata rubicola*). *The Auk* **119**(4), 1167-1173 (2002).
- Washburn, B.E., Morris, D.L., Millspaugh, J.J., et al. Using a commercially available radioimmunoassay to quantify corticosterone in avian plasma. *The Condor* **104**, 558-63 (2002).
- Evans, M.R., Roberts, M.L., Buchanan, K.L., et al. Heritability of corticosterone response and changes in life history traits during selection in the zebra finch. *J. Evol. Biol.* **19**(2), 343-52 (2005).
- Martínex-Cué, C., Rueda, N., Garcia, E., et al. Anxiety and panic responses to predator in male and female Ts65Dn mice, a model for Down syndrome. *Genes Brain Behav.* **5**, 413-422 (2006).
- Meijer, M.K., Lemmens, A.G., and Van Zutphen, B.F.M. Urinary corticosterone levels in mice in response to intraperitoneal injections with saline. *J. Appl. Anim. Welf. Sci.* **8**(4), 279-283 (2005).
- Möstl, E., Rettenbacher, S., and Palme, R. Measurement of corticosterone metabolites in birds' droppings: An analytical approach. *Ann. N. Y. Acad. Sci.* **1046**, 17-34 (2005).
- Bortolotti, G.R., Marchant, T.A., Blas, J., et al. Corticosterone in feathers is a long-term, integrated measure of avian stress physiology. *Funct. Ecol.* **22**, 494-500 (2008).
- Vázquez-Palacios, G., Retana-Márquez, S., Bonilla-Jaime, H., et al. Further definition of the effect of corticosterone on the sleep-wake pattern in the male rat. *Pharmacol. Biochem. Behav.* **70**, 305-310 (2001).
- Miller, W.L. and Tyrrell, J.B. The adrenal cortex, Chapter 12, in *Endocrinology and Metabolism*. Felig, P., Baxter, J.D., and Frohman, L.A., editors, 3rd edition, McGraw-Hill, Inc. Health Professions Division, New York, 555-711 (1995).

11. Maclouf, J., Grassi, J., and Pradelles, P. Development of enzyme-immunoassay techniques for the measurement of eicosanoids, Chapter 5, *in Prostaglandin and Lipid Metabolism in Radiation Injury*. Walden, T.L., Jr. and Hughes, H.N., editors, Plenum Press, Rockville, 355-364 (1987).
12. Pradelles, P., Grassi, J. and Maclouf, J. Enzyme immunoassays of eicosanoids using acetylcholinesterase as label: An alternative to radioimmunoassay. *Anal. Chem.* **57**, 1170-1173 (1985).
13. Maxey, K.M., Maddipati, K.R. and Birkmeier, J. Interference in enzyme immunoassays. *J. Clin. Immunoassay* **15**,116-120 (1992).

12									
11									
10									
9									
8									
7									
6									
5									
4									
3									
2									
1									
	A	B	C	D	E	F	G	H	

NOTES

Warranty and Limitation of Remedy

Buyer agrees to purchase the material subject to Cayman's Terms and Conditions. Complete Terms and Conditions including Warranty and Limitation of Liability information can be found on our website.

This document is copyrighted. All rights are reserved. This document may not, in whole or part, be copied, photocopied, reproduced, translated, or reduced to any electronic medium or machine-readable form without prior consent, in writing, from Cayman Chemical Company.

©01/29/2015, Cayman Chemical Company, Ann Arbor, MI, All rights reserved.
Printed in U.S.A.