
Product Manual


OxiSelect™ Intracellular ROS Assay Kit (Green Fluorescence)

Catalog Number

STA-342

96 assays

FOR RESEARCH USE ONLY
Not for use in diagnostic procedures



CELL BIOLABS, INC.
Creating Solutions for Life Science Research

Introduction

Accumulation of reactive oxygen species (ROS) coupled with an increase in oxidative stress has been implicated in the pathogenesis of several disease states. The role of oxidative stress in vascular diseases, diabetes, renal ischemia, atherosclerosis, pulmonary pathological states, inflammatory diseases, and cancer has been well established. Free radicals and other reactive species are constantly generated *in vivo* and cause oxidative damage to biomolecules, a process held in check by the existence of multiple antioxidant and repair systems as well as the replacement of damaged nucleic acids, proteins and lipids. Measuring the effect of antioxidant therapies and ROS activity intracellularly is crucial to suppressing or treating oxidative stress inducers.

Cell Biolabs' OxiSelect™ Intracellular ROS Assay Kit (Green Fluorescence) is a cell-based assay for measuring hydroxyl, peroxy, or other reactive oxygen species activity within a cell. The assay employs the cell-permeable fluorogenic probe 2', 7'-Dichlorodihydrofluorescein diacetate (DCFH-DA). In brief, DCFH-DA is diffused into cells and is deacetylated by cellular esterases to non-fluorescent 2', 7'-Dichlorodihydrofluorescein (DCFH), which is rapidly oxidized to highly fluorescent 2', 7'-Dichlorodihydrofluorescein (DCF) by ROS (Figure 1). The fluorescence intensity is proportional to the ROS levels within the cell cytosol. The effect of antioxidant or free radical compounds on DCFH-DA can be measured against the fluorescence of the provided DCF standard. The kit has a DCF detection sensitivity limit of 10 pM. Each kit provides sufficient reagents to perform up to 96 assays, including standard curve and unknown samples.

Assay Principle

The OxiSelect™ Intracellular ROS Assay Kit is a cell-based assay for measuring antioxidant or ROS activity. Cells are cultured in a 96-well cell culture plate and then pre-incubated with DCFH-DA, which is cell-permeable. The unknown antioxidant or ROS samples are then added to the cells. After a brief incubation, the cells can be read on a standard fluorescence plate reader. The ROS or antioxidant content in unknown samples is determined by comparison with the predetermined DCF standard curve.

Related Products

1. STA-320: OxiSelect™ Oxidative DNA Damage ELISA Kit (8-OHdG Quantitation)
2. STA-330: OxiSelect™ TBARS Assay Kit (MDA Quantitation)
3. STA-332: OxiSelect™ MDA Adduct ELISA Kit
4. STA-334: OxiSelect™ HNE-His Adduct ELISA Kit
5. STA-341: OxiSelect™ Catalase Activity Assay Kit
6. STA-343: OxiSelect™ Hydrogen Peroxide Assay Kit
7. STA-344: OxiSelect™ Hydrogen Peroxide/Peroxidase Assay Kit
8. STA-345: OxiSelect™ ORAC Activity Assay Kit
9. STA-347: OxiSelect™ In Vitro ROS/RNS Assay Kit (Green Fluorescence)
10. STA-350: OxiSelect™ Comet Assay Kit (3-Well Slides), 15 assays




Figure 1. Mechanism of DCF Assay

Kit Components

1. 20X DCFH-DA (Part No. 234201): One 500 μ L amber tube of a 20 mM solution in methanol.
2. DCF Standard (Part No. 234202): One 100 μ L amber tube of a 1 mM solution in DMSO.
3. Hydrogen Peroxide (Part No. 234102): One 100 μ L amber tube of an 8.821 M solution.
4. 2X Cell Lysis Buffer (Part No. 234203): One 20 mL bottle.

Materials Not Supplied

1. Sterile DPBS for washes and buffer dilutions
2. Hank's Balanced Salt Solution (HBSS)
3. Cell culture medium (ie: DMEM +/-10% FBS)
4. 96-well black or fluorescence microtiter plate
5. Fluorescent microplate reader capable of reading 480 nm (excitation) and 530 nm (emission)

Storage

Upon receipt, store the DCFH-DA and DCF Standard at -20°C. Avoid multiple freeze/thaw cycles. Store the Cell Lysis Buffer and Hydrogen Peroxide at 4°C.

Preparation of Reagents

- 1X DCFH-DA: Dilute the 20X DCFH-DA stock solution to 1X in cell culture media, preferably without FBS. Stir or vortex to homogeneity. Prepare only enough for immediate applications.

Notes:

- *1X DCFH-DA/media solution contains 5% methanol. For cells that are sensitive to methanol, we recommend instead preparing a 0.1X (100 µM) solution of DCFH-DA in cell culture media.*
- *Due to light-induced auto-oxidation, DCFH-DA solutions at any concentration must be protected from light.*
- Hydrogen Peroxide (H₂O₂): Prepare H₂O₂ dilutions in DMEM or DPBS as necessary. Do not store diluted solutions.

Preparation of Standard Curve

1. Prepare a 1:10 dilution series of DCF standards in the concentration range of 0 µM – 10 µM by diluting the 1 mM DCF stock in cell culture media (see Table 1).

Standard Tubes	DCF Standard (µL)	Culture Medium (µL)	DCF (nM)
1	10	990	10,000
2	100 of Tube #1	900	1000
3	100 of Tube #2	900	100
4	100 of Tube #3	900	10
5	100 of Tube #4	900	1
6	100 of Tube #5	900	0.1
7	100 of Tube #6	900	0.01
8	0	1000	0

Table 1. Preparation of DCF Standards

2. Transfer 75 μ L of each DCF standard to a 96-well plate suitable for fluorescence measurement. Add 75 μ L of the 2X Cell Lysis Buffer.
3. Read the fluorescence with a fluorescence plate reader at 480 nm excitation /530 nm emission.

Assay Protocol

I. DCF Dye Loading

1. Prepare and mix all reagents thoroughly before use. Each cell sample, including unknown and standard, should be assayed in duplicate or triplicate.
2. Culture cells in either a clear or black 96-well cell culture plate.
Note: If using a black plate, choose an appropriate plate based on your fluorometer's reader (i.e. choose a clear bottom black plate for bottom readers).
3. Remove media from all wells and discard. Wash cells gently with DPBS or HBSS 2-3 times. Remove the last wash and discard.
4. Add 100 μ L of 1X DCFH-DA/media solution to cells. Incubate at 37°C for 30-60 minutes.
5. Remove solution. Repeat step three using multiple washes with DPBS or HBSS. Remove the last wash and discard.
6. Treat DCFH-DA loaded cells with desired oxidant or antioxidant in 100 μ L medium.

II. Quantitation of Fluorescence

- Fluorescence microscopy or Flow cytometry: Fluorescence can be analyzed on an inverted fluorescence microscope or by flow cytometry using excitation and emission wavelengths of 480 nm and 530 nm, respectively.
- Fluorescence Plate Reader:
 - Assays performed in black cell culture fluorometric plates: Plate may be read immediately for kinetic analysis or after 1 hour for static analysis. Plates read for kinetic analysis may be read in increments of 1 and 5 minutes up to 1 hour or more as necessary. Read the fluorescence with a fluorometric plate reader at 480 nm/530 nm.
 - Assays performed in clear cell culture plates: After treatment with desired oxidant or antioxidant, carefully remove treatment media from all wells and discard. Wash cells gently with DPBS or HBSS 2-3 times. Remove the last wash and discard. Add 100 μ L of medium to the wells and terminate the assay by adding 100 μ L of the 2X Cell Lysis Buffer. Mix thoroughly and incubate 5 minutes. Transfer 150 μ L of the mixture to a 96-well plate suitable for fluorescence measurement. Read the fluorescence with a fluorometric plate reader at 480 nm/530 nm.

Example of Results

The following figures demonstrate typical ROS Assay results. Fluorescence measurement was performed on SpectraMax Gemini XS Fluorometer (Molecular Devices) with a 485/538 nm filter set and 530 nm cutoff. One should use the data below for reference only. This data should not be used to interpret actual results.




Figure 2. DCF Standard Curve.




Figure 3. ROS in HeLa cells treated with H₂O₂. 50,000 HeLa cells in a 96-well plate were first pretreated with 1 mM DCFH-DA for 60 minutes at 37°C. Cells were then treated with various concentrations of H₂O₂ for 20 minutes.




Figure 4. DCF Fluorescence in H₂O₂ treated HeLa cells after 1 hour. Left: 0 μM; Right: 1000 μM.

References

1. Bass DA, Parce JW, Dechatelet LR, Szejda P, Seeds MC, Thomas M. Flow cytometric studies of oxidative product formation by neutrophils: A graded response to membrane stimulation. *J Immunol.* 1983; 130:1910-1917.
2. Brandt R, Keston AS. Synthesis of diacetyldichlorofluorescein: A stable reagent for fluorometric analysis. *Anal Biochem.* 1965; 11:6-9.
3. Keston AS, Brandt R. The fluorometric analysis of ultramicro quantities of hydrogen peroxide. *Anal Biochem.* 1965; 11:1-5.

Recent Product Citations

1. Cho, K.A. et al. (2012). IL-17 and IL-22 Enhance Skin Inflammation by Stimulating the Secretion of IL-1β by Keratinocytes via the ROS-NLRP3-Caspase-1 Pathway. *Int. Immunol.* **24**:147-158.
2. Martin, S.A. et al. (2011). Parallel High-Throughput RNA Interference Screens Identify PINK1 as a Potential Therapeutic Target for the Treatment of DNA Mismatch Repair-Deficient Cancers. *Cancer Res.* **71**:1836-1848.
3. Kuznetsov, J.N. et al. (2011). AMPK and Akt Determine Apoptotic Cell Death following Perturbations of One-Carbon Metabolism by Regulating ER Stress in Acute Lymphoblastic Leukemia. *Mol. Cancer Ther.* **10**:437-447.
4. Wei, J. et al. (2011). C-Jun N-terminal Kinase (JNK-1) Confers Protection Against Brief but not Extended Ischemia during Acute Myocardial Infarction. *J. Biol. Chem.* **286**:13995-14006.
5. Titos, E. et al. (2011). Resolvin D1 and its Precursor Docosahexaenoic Acid Promote Resolution of Adipose Tissue Inflammation by Eliciting Macrophage Polarization toward an M2-Like Phenotype. *J. Immunol.* **187**:5408-5418.
6. Zhang, J. et al. (2011). 2-Deoxy-D-Glucose Attenuates Isoflurane-Induced Cytotoxicity in an In Vitro Cell Culture Model of H4 Human Neuroglioma Cells. *Anesth. Analg.* **113**:1468-1475.
7. Zhang, Y. et al. (2010). The Mitochondrial Pathway of Anesthetic Isoflurane-Induced Apoptosis. *J. Biol. Chem.* **285**:4025-4037.
8. Batova, A. et al. (2010). The Synthetic Caged *Garcinia* Xanthone Cluvenone Induces Cell Stress and Apoptosis and Has Immune Modulatory Activity. *Mol. Cancer Ther.* **9**:2869-2878.
9. Tanaka, N. et al. (2010). *Cis*-dichlorodiammineplatinum Upregulates Angiotensin II Type 1 Receptors through Reactive Oxygen Species Generation and Enhances VEGF Production in Bladder Cancer. *Mol. Cancer Ther.* **9**:2982-2992

Warranty

These products are warranted to perform as described in their labeling and in Cell Biolabs literature when used in accordance with their instructions. THERE ARE NO WARRANTIES THAT EXTEND BEYOND THIS EXPRESSED WARRANTY AND CELL BIOLABS DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR WARRANTY OF FITNESS FOR PARTICULAR PURPOSE. CELL BIOLABS' sole obligation and purchaser's exclusive remedy for breach of this warranty shall be, at the option of CELL BIOLABS, to repair or replace the products. In no event shall CELL BIOLABS be liable for any proximate, incidental or consequential damages in connection with the products.

Contact Information

Cell Biolabs, Inc.
7758 Arjons Drive
San Diego, CA 92126
Worldwide: +1 858-271-6500
USA Toll-Free: 1-888-CBL-0505
E-mail: tech@cellbiolabs.com
www.cellbiolabs.com

©2009-2012: Cell Biolabs, Inc. - All rights reserved. No part of these works may be reproduced in any form without permissions in writing.